

	<div>TECHNICAL FILE - EU DECLARATION OF CONFORMITY (MDR)</div>	<div>Réf : FQ279B Date : 21/07/22</div>

We

Name + address + Single Registration Number of manufacturer:	Laboratoires Anios 1 rue de l'Espoir 59260 Lezennes France FR-MF-000011756
Name + address + Single Registration Number of authorized representative:	Not Applicable

declare on our sole responsibility that

the medical device	ANIOS LB 30
Type / Intended Use	<i>RINSING AGENT INVASIVE / Acid rinsing for thermal wash basins</i>
Class	<i>I</i>
Rule according to MDR Annex VIII	<i>Rule 1</i>

Meets all the provisions of the Regulation (EU) 2017/745 on medical devices.

Notified body name, address, ID	Not Applicable
Conformity assessment procedure	<i>MDR (EU) 2017/745 article 52, Annex II and III</i>
ID of the certificates issued	Not Applicable
Common Specifications	Not Applicable
Validity	26 May 2024

Valid in conjunction with the batch related release documentation

Sainghin 09/01/2022	Dieter Wirbals	Sainghin 09/01/2022	Jean-Noël Bertho
Place, date	Name and function (Person Responsible for Regulatory Compliance)	Place, date	Name and function (RD,E Representative)

Signed on behalf of: **Laboratoires Anios**

Medical Device	Brand	Device Subcategory	Packaging variant	Basic UDI-DI
ANIOS LB 30	ANIOS	Rinsing agent invasive	38 2x5L	359761LA01501NLIN3