

	TECHNICAL FILE - EU DECLARATION OF CONFORMITY (MDR)	<b>Réf : FQ279B</b> <b>Date : 21/07/22</b>

We

Name + address + Single Registration Number of manufacturer:	<b>Laboratoires Anios</b> <b>1 rue de l'Espoir</b> <b>59260 Lezennes</b> <b>France</b> <b>FR-MF-000011756</b>
Name + address + Single Registration Number of authorized representative:	Not Applicable

declare on our sole responsibility that

the medical device	ANIOSYME FOAM
Type / Intended Use	<i>DETERGENT INVASIVE / Instrument cleaning</i>
Class	<i>I</i>
Rule according to MDR Annex VIII	<i>Rule 1</i>

Meets all the provisions of the Regulation (EU) 2017/745 on medical devices.

Notified body name, address, ID	Not Applicable
Conformity assessment procedure	<i>MDR (EU) 2017/745 article 52, Annex II and III</i>
ID of the certificates issued	Not Applicable
Common Specifications	Not Applicable
Validity	26 May 2024

Valid in conjunction with the batch related release documentation

Sainghin 09/01/2023		Sainghin 09/01/2023	
	Dieter Wirbals		Jean-Noël Bertho
Place, date	Name and function (Person Responsible for Regulatory Compliance)	Place, date	Name and function (RD,E Representative)

Signed on behalf of: **Laboratoires Anios**

Medical Device	Brand	Device Subcategory	Packaging variant	Basic UDI-DI
ANIOSYME FOAM	ANIOS	Detergent invasive	544 12x750mL	359761LA0150C1LIME